

July 2015

Risk of blood loss from a haemodialysis line caused by leakage at the blood chamber connection

It has been reported by the MHRA that a manufacturer of a haemodialysis line has reported that there is a risk of blood loss caused by leakage at the blood chamber connection.

<https://www.gov.uk/drug-device-alerts/haemodialysis-crit-line-blood-chamber-with-product-code-cl10021021-recall-due-to-risk-of-blood-loss>

Action

- Identify and quarantine affected devices listed in the Field Safety Notice.
- Return the reply form
- Contact the manufacturer via email for advice on how to return affected product to the manufacturer.

Please submit comments, solutions, and personal experience to:

Dr. Paul Rylance, RA-BRS Patient Safety lead by email to: paul.rylance@nhs.net